

CS-42

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name D:\Data\SongOct2015\CH-216.d
Method lowpositive.m
Sample Name PA
Comment

Acquisition Date 10/1/2015 4:01:02 PM

Operator Lijiang Song
Instrument maXis plus 255552.00050

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.5 Bar
Focus	Not active	Set Capillary	4500 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	2400 m/z	Set Charging Voltage	0 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C



CH-216.d

Bruker Compass DataAnalysis 4.2

printed: 10/1/2015 4:05:19 PM

by: Lijiang Song

Page 1 of 2