

LSJ-43

## Mass Spectrum SmartFormula Report

### Analysis Info


Analysis Name D:\Data\SongSept2015\CH218.d  
Method lowpositive.m  
Sample Name LS  
Comment

Acquisition Date 9/30/2015 3:49:57 PM

Operator Lijiang Song  
Instrument maXis plus 255552.00050

### Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.5 Bar
Focus	Not active	Set Capillary	4500 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	2400 m/z	Set Charging Voltage	0 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C



CH218.d

Bruker Compass DataAnalysis 4.2

printed: 9/30/2015 3:53:02 PM

by: Lijiang Song

Page 1 of 2