

157-44

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name D:\Data\SongSept2015\CH217.d
Method lowpositive.m
Sample Name LS
Comment

Acquisition Date 9/30/2015 3:48:51 PM

Operator Lijiang Song
Instrument maXis plus 255552.00050

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.5 Bar
Focus	Not active	Set Capillary	4500 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	2400 m/z	Set Charging Voltage	0 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C



CH217.d

Bruker Compass DataAnalysis 4.2

printed: 9/30/2015 3:53:44 PM

by: Lijiang Song

Page 1 of 2