

University of Warwick institutional repository: <http://go.warwick.ac.uk/wrap>

This paper is made available online in accordance with publisher policies. Please scroll down to view the document itself. Please refer to the repository record for this item and our policy information available from the repository home page for further information.

To see the final version of this paper please visit the publisher's website. Access to the published version may require a subscription.

Author(s): Neil Stewart

Article Title: Millisecond accuracy video display using OpenGL under Linux

Year of publication: 2006

Link to published version: <http://brm.psychonomic-journals.org/content/38/1/142.full.pdf+html>

Publisher statement: None

Running Head: LINUX VIDEO DISPLAY

Millisecond Accuracy Video Display Using OpenGL Under Linux

Neil Stewart

University of Warwick, England

Stewart, N. (2006). Millisecond accuracy video display using OpenGL under Linux. *Behavior Research Methods*, 38, 142-145.

Neil Stewart

Department of Psychology

University of Warwick

Coventry

CV4 7AL

England

+44 24 7657 3127

neil.stewart@warwick.ac.uk

Abstract

To measure people's reaction times to the nearest millisecond, it is necessary to know exactly when a stimulus is displayed. This article describes how to display stimuli with millisecond accuracy on a normal CRT monitor using a PC running Linux. A simple C program is presented to illustrate how this may be done within X Windows using the OpenGL rendering system. A test of this system is reported that demonstrates that stimuli may be consistently displayed with millisecond accuracy. An algorithm is presented that allows the exact time of stimulus presentation to be deduced, even if there are relatively large errors in measuring the display time.

Millisecond Accuracy Video Display Using OpenGL Under Linux

In many psychology experiments, a stimulus is presented and the latency of a participant's response is recorded. There are two key pieces of information that are required to determine a participant's reaction time: the time that the stimulus was presented and the time that the participant responded. I describe how to record responses with millisecond accuracy under Linux in Stewart (2004). This article is concerned with presenting stimuli with millisecond accuracy.

A Choice of Display Method

A standard cathode ray tube (CRT) monitor works by scanning a beam of electrons over a phosphorus-coated screen. When the beam strikes the phosphorus, the phosphorus fluoresces. The beam scans each row of pixels from left to right, beginning with the top row and ending with the bottom row. The intensity of the beam is varied with location, to cause the phosphorus to fluoresce with different intensity, allowing a picture to be built up over a scan of the whole screen. The number of scans of the entire screen per second is called the refresh rate.

Recently, liquid crystal displays (LCDs) have been introduced. These work quite differently. When a voltage is applied to a liquid crystal, it changes its shape. Different shaped crystals rotate the plane of polarized light by different extents. In an LCD, a layer of liquid crystals is sandwiched between two orthogonal polarizers. Unless the plane of the polarized light is rotated in between the two polarizers it cannot pass through them. By varying the voltage applied to the liquid crystals in the sandwich, the degree of rotation of the plane of the polarized light varies, allowing more or less light to pass through the LCD. By having a large array of liquid crystal segments, each controlled independently, the amount of light passing through different parts of the LCD can be controlled to create a picture. In thin film transistor (TFT) LCD panels, a transistor and capacitor assembly for each pixel is used to maintain the voltage across the liquid crystal between refreshes of the display.

Plant, Hammond, and Turner (2004) studied the performance of some CRT displays and some TFT LCD displays, by measuring the time for the display to become visible (as measured by a photodiode and oscilloscope apparatus). The CRTs tested reached full brightness in less than 1 ms. However, the brightness of the LCDs tested changed very slowly, taking 10 - 20 ms for the display to reach full brightness (i.e., to warm up), depending on the display. Further, the LCD displays took 5 - 20 ms for the image to fade to the background level of brightness on stimulus offset (i.e., to cool down).¹ The point during the warm-up time at which an image is detectable by a human observer and similarly the point when it is no longer visible during the cool-down time is likely to vary with environmental factors (e.g., ambient illumination, the brightness of other stimuli) and history (e.g., the brightness of recent stimuli). Thus, Plant et al.'s study shows that LCD TFT displays are not suitable for presenting stimuli with millisecond accuracy. Here I describe how to control a CRT display with millisecond accuracy in Linux.

A Simple Linux Program²

The most commonly used graphics system under Linux is X Windows. X Windows provides a graphical engine that is hardware independent and used across a large variety of operating systems such as Mac OS X and Sun Solaris. Linux systems usually use a free implementation of the X Windows standard called XFree86 (<http://www.xfree86.org/>). The X Windows system performs windowing tasks, such as allocating new windows, moving them around, and deciding when they need to be redrawn. In this simple program I have used the OpenGL graphics library to draw into a new window that fills the entire screen. OpenGL is also hardware independent and runs on many other operating systems. I used the Mesa3D implementation of the OpenGL standard (<http://www.mesa3d.org/>).³

The complete C source code for the program is available from <http://www.warwick.ac.uk/staff/Neil.Stewart/>. The program first creates a blank, borderless window that fills the entire screen. This is done using X and OpenGL library functions. If you

would like more information than is provided by the man pages for these functions, consult Nye (1992) and Woo, Neider, Davis, & Shreiner (1999). First, `XOpenDisplay` is used to open a connection to the X Server and obtain server *display* information. Second, `glXChooseVisual` is used to select, from all available *visuals* on the display, the type of visual that best meets a minimum specification. Third, `glXCreateContext` is used to create a *rendering context* with direct access to the video card. Fourth, a new window that fills the entire screen and has no border is made using `XCreateWindow`. The attributes for this window are extracted from those of the root window. Fifth, to display or *map* the window to the screen `XMapWindow` is called followed by `XIfEvent`. `XIfEvent` waits for the window to be displayed before allowing execution to proceed.

So far, a new window has been created with an OpenGL rendering context that renders directly to the video card. Next, the details of the OpenGL system need to be specified. The rendering context is selected with `glXMakeCurrent`, and the coordinate system established with `glViewport`. OpenGL, which is a 3D rendering system, can be set to 2D by setting the model and projection transforms to the identity matrix (by calling `glMatrixMode` and `glLoadIdentity`) and calling `glOrtho2D`.

The final step is to hide the cursor. Unfortunately, to hide the cursor X Windows requires the creation of a blank cursor using `XCreatePixmapCursor` by passing a blank pixmap created with `XCreateBitmapFromData` as a parameter. The cursor is defined as the current cursor using `XDefineCursor`. Together, all of these steps may seem rather lengthy, but they will work on any X Windows System with OpenGL.

From this point, things are much more simple. The rendering context that the program selects is a double buffered context. In double buffering, while the contents of one buffer are being displayed, drawing takes place on the other buffer. When a `glXSwapBuffers` call is made, the roles of the two buffers are reversed. The reversal is synchronized to the screen retrace, so that the buffers are swapped while the beam is moving from the bottom right

corner at the end of a trace back to the top left corner for the beginning of the next. This means that, if buffers are swapped every frame, there is one frame period (e.g., 10 ms at 100 Hz refresh rate) for drawing into a buffer. Because modern computers are very fast, it is possible to draw an awful lot in this time (cf. current computer games like Doom 3 and Half-Life 2) depending on the processor and graphics card. There is no longer a need to buffer a really large number of frames in advance (cf. only 5 years ago, Graves & Bradley, 1988a, 1988b; Myors, 1998, 1999; Tsai, 2001).

As an example, the following code fragment draws a red rectangle in the center of a green screen.

```
glClearColor(0.0, 1.0, 0.0, 1.0);
glClear(GL_COLOR_BUFFER_BIT);
glColor3f(1.0, 0.0, 0.0);
glRecti(-100, -100, 100, 100);
glXSwapBuffers(display, window);
glFinish();
gettimeofday(time, (struct timezone *) 0);
```

The clear color is set with `glClearColor` (which takes red, green, blue, and alpha values) before the screen is cleared by `glClear`. The drawing color is set to red by `glColor3f` (which takes red, green, and blue values) and a rectangle is drawn with `glRecti` with the bottom left corner at -100, -100 and the top right corner at 100, 100. The call to `glXSwapBuffers` swaps the buffers, displaying the rectangle. The `glFinish` call ensures that all commands are complete before it returns. Thus the call to `gettimeofday` sets the variable *time* to the time at which the tracing of first frame containing the rectangle began.

A Test of Millisecond Accuracy

I conducted a simple test of the accuracy of the times recorded as above. I set up a simple loop to alternate the screen colors on every retrace of the screen between black and

white, and recorded the time after every call to `glXSwapBuffers` and `glFinish`. If the times are accurate, then the intervals between times should be exactly the refresh period.

The system I ran these tests on is a dual processor 1.4 MHz AMD Athlon with 256 MB of RAM with a PCI NVidia GeForce 2 MX with 32 MB of video RAM. I used the Debian Woody GNU distribution with a 2.4.25-1-k7-smp Linux kernel, XFree86 4.1.0 and the NVidia 1.0-5336 driver⁴ (and not the `nv` driver that comes with XFree86). The CRT monitor was a Sony CPD-G220 Color Trinitron running with a refresh rate of 85 Hz. The following source code was used to sample the time of 10,000 consecutive refreshes.

```
for(i=0; i<10000; i++) {
    c=double(i%2);
    glClearColor(c, c, c, 1.0);
    glClear(GL_COLOR_BUFFER_BIT);
    glXSwapBuffers(display, window);
    glFinish();
    gettimeofday(time[i], (struct timezone *) 0);
}
```

By alternating the color of the screen on alternate refreshes between black and white (using the variable `c`), a photodiode apparatus that was connected to the screen could detect the refreshes. During the test, no refreshes were missed (i.e., the screen color swapped every 1/85 s).

For a vertical refresh rate of 85 Hz, the period should be 11.765 ms. Figure 1 shows the differences between the consecutive times recorded in the array `time`. Only 12 intervals (about 1 in 1000) fell outside that range ± 0.05 ms from the expected value of 11.765 ms. These cases occurred on the six occasions when there was a large gap between the calls to `glXSwapBuffers` and `glFinish` and the call to `gettimeofday`. The large gap caused the preceding inter-refresh interval to be overestimated and the following inter-refresh interval to be

underestimated by the same amount. The largest error was 5 ms. The very rare errors of up to 5 ms may not be of any concern, because they are so rare, but if they are, fortunately, they are very easy to correct using the algorithm described below.

A second, more realistic test was also run. The duration of a white screen presented for 1000 ms (i.e., 85 retraces at an 85 Hz refresh rate) was measured for each of 1000 trials. The inter-trial interval was also 1000 ms, during which time a black screen was presented. The total duration of the test was about half an hour. Across the 1000, trials the minimum recorded display time was 999.734 ms and the maximum was 999.800 ms. This gives a total range of only 0.066 ms for the stimulus durations, which is an accuracy of better than one tenth of a ms.⁵

A Simple Correction

For a monitor driven at ν Hz, the times of the beginnings of the refresh must all be $1/\nu$ s apart. By taking advantage of this fact, the exact time of stimulus retraces can be deduced, providing the error in recording them is small compared to the refresh period $1/\nu$. This model can be formalized in Equation 1.

$$t_i = t_0 + i / \nu \quad (1)$$

where t_i is the time (in s) of the i th retrace and i indexes the number of retraces since the first at time t_0 . Suppose that one has recorded a series of retrace times r_j (where j indexes the number of recordings since t_0) and that there is some noise in the recording that varies from trial to trial, so that

$$r_j = t_i + \epsilon_j \quad (2)$$

where $\epsilon_j > 0$ is the error in recording the time. From Equations 1 and 2, $r_j - t_0 = i / \nu + \epsilon_j$. In other words, apart from the error in recording the time, $r_j - t_0$ must be an integer multiple of the refresh period $1/\nu$. Under the assumption that $\epsilon_j < 1/\nu$, then

$$i = \text{Quotient}[r_j - t_0, 1 / \nu] \quad (3)$$

where the function $\text{Quotient}[\dots]$ gives the integer part of the division of the first parameter

by the second. That is, the retrace i on which the j th recording r_j was made can be determined. Once i is known, the exact time of the retrace i is given by Equation 1.

t_0 is, however, not known exactly. But it can be estimated from the data. First, construct the error term

$$error = (\text{Mod}[r_j - t_0, 1 / v])^2 \quad (4)$$

where the function $\text{Mod}[\cdot, \cdot]$ returns the remainder after the first parameter is divided by the second. Error is the squared deviation between the recorded time r_j and the first preceding modeled retrace time (the location of which is determined by t_0). Second, average this term over all of the refresh times recorded. Third, find the value of t_0 for which this sum is a minimum.

To test this algorithm, I have generated 1000 trials of artificial refresh data using Equations 1 and 2 as follows. The retraces on which recordings were made (i.e., the i s) were generated by starting with a seed value of $i = 0$, and generating each subsequent value by adding to the previous value an integer drawn at random from a uniform distribution of integers between 1 and 20. In this way, the retraces recorded were irregularly spaced, as they might be in an experiment where the inter-stimulus intervals depend on a participant's response times. The noise ϵ_j was modeled as a log normal distribution with mean 2 ms and standard deviation 2 ms. This standard deviation is 20 times greater than the variability observed in the above practical test, and so this test is comparatively very tough. t_0 was set to an arbitrary value of 325 ms. In this way, 1000 simulated r_j s were generated.

The average error (Equation 4) is plotted as a function of t_0 in Figure 2. The function is periodic, and has a minimum every $1 / v$ s. The minima immediately preceding r_0 is t_0 . Once t_0 is known, Equation 3 can be used to recover the numbers of the retraces (i.e., the i s) upon which the recordings were made. Substituting these i s into Equation 1 allows the exact retrace times to be calculated. As a check, the distribution of residuals ($r_j - t_i$) can be plotted. Figure 3A shows the distribution with $t_0 = 325$. However, if an incorrect value for t_0 is used, the

distribution becomes bimodal. Figure 3B illustrates this occurring when $t_0 = 315$.

In summary, by taking advantage of the fact that retraces can only occur at certain predefined times, noisy timing data can be cleaned up, providing a much better estimate of the time at which a stimulus is displayed and eliminating trial-to-trial variability. Although the Linux system described above almost always performed very accurately, this method allows the very rare errors to be corrected.

Conclusion

In conclusion, a PC Linux system and a CRT monitor can be used to display stimuli with millisecond accuracy.

References

- Graves, R., & Bradley, R. (1988a). More on millisecond timing and tachistoscope applications for the IBM PC. *Behavior Research Methods, Instruments, & Computers*, 20, 408-412.
- Graves, R., & Bradley, R. (1988b). Millisecond timing on the IBM PC/XT/AT and PS/2: A review of the options and correction for the Graves and Bradley algorithm. *Behavior Research Methods, Instruments, & Computers*, 23, 377-379.
- Finney, S. A. (2001). Real-time data collection in Linux: A case study. *Behavior Research Methods, Instruments, and Computers*, 33, 167-173.
- MacInnes, W. J., & Taylor, T. L. (2001). Millisecond timing on PCs and Macs. *Behavior Research Methods, Instruments, & Computers*, 33, 174-178.
- Myors, B. (1998a). The PC tachistoscope has 32 pages. *Behavior Research Methods, Instruments, & Computers*, 30, 457-461.
- Myors, B. (1999). The PC tachistoscope has 240 pages. *Behavior Research Methods, Instruments, & Computers*, 31, 329-333.
- Nye, A. (1992). *Xlib programming manual* (3rd ed.). Sebastopol, CA: O'Reilly.
- Plant, R. P., Hammond, N., & Turner, G. (2004). Self-validating presentation and response timing in cognitive paradigms: How and why. *Behavior, Research Methods, Instruments, & Computers*, 36, 291-303.
- Stewart, N. (2004). *A PC parallel port button box provides millisecond response time accuracy under Linux*. Manuscript submitted for publication.
- Tsai, J. L. (2001). A multichannel PC tachistoscope with high resolution and fast display change capability. *Behavioral Research Methods, Instruments, & Computers*, 33, 524-531.
- Woo, M., Neider, J., Davis, T., & Shreiner, D. (1999). *OpenGL programming guide* (3rd ed.). Reading, MA: Addison Wesley.

Author Note

Neil Stewart, Department of Psychology University of Warwick.

I thank Dave Sleight for constructing the photodiode apparatus and James Adelman and Matthew Roberts for sharing their Linux knowledge.

Correspondence concerning this article should be addressed to Neil Stewart, Department of Psychology, University of Warwick, Coventry, CV4 7AL, UK. E-mail: neil.stewart@warwick.ac.uk.

Footnotes

¹Plant et al. (2003) also found that there was a delay of up to 50 ms between the time that E-prime software was scheduled to display the image and the time the image began to appear on the display.

²For a multitasking operating system (e.g., Linux, MacOS X, or MS Windows) to be able to function with millisecond accuracy, one must prevent the program from being interrupted and having its memory swapped to disc. An accurate timer is also required. Finney (2001; see also MacInnes & Taylor, 2001; Stewart, 2004) describes how to ensure a Linux program meets this standard.

³The OpenGL utility toolkit (GLUT) can be used to simplify the process of initializing and opening new windows and controlling window events. In GLUT, buffers are swapped by calling `glutSwapBuffers`, which in turn calls `glXSwapBuffers`.

⁴To ensure that buffers are swapped only during the vertical blanking period by `glXSwapBuffers`, the environment variable `__GL_SYNC_TO_VBLANK` must be set to a non-zero value. It is easy to spot when buffer swapping is not synchronized, because the screen seems to "tear".

⁵The fact that the total display time was slightly less than 1000 ms indicates either that the actual monitor refresh rate was slightly faster than 85 Hz or that the computer's clock is running slightly slow. The clock on the test computer loses about 2.6×10^{-7} s per s, as measured by the network time protocol software (<http://www.ntp.org/>).

Figure Caption

Figure 1. The distribution of inter-refresh intervals measured by a `gettimeofday` call after a `glXSwapBuffers` and a `glFinish`.

Figure 2. Model error as a function of t_0 .

Figure 3. The distribution of residuals with t_0 correct (A) and incorrect (B).

Figure 1




Figure 2





Figure 3

A



B

